

R5 - LAVAGE CONCENTRE **CE** CONCENTRATED WASH BUFFER



Réactif pour immunoblot de diagnostic *in vitro*
Technique manuelle / semi-automatisable

Reagent for *In vitro* diagnostic Immunoblot assay
Semi-automated / manual technique

#WB-LC60
#WB-LC250

NOTICE D'UTILISATION INSTRUCTIONS FOR USE – page 3

Retrouvez plus d'informations et les notices traduites dans votre langue
sur notre site internet www.ldbiodiagnostics.com

Find more information and IFU in your language on our website www.ldbiodiagnostics.com

INDICATION DU TEST

Se référer à la notice du kit LDBIO Diagnostics.
Ce réactif ne doit pas être utilisé seul.

PRINCIPE DU TEST / DEROULEMENT DU TEST

Technique de Western Blot : Le tampon de lavage est utilisé pour le lavage des bandelettes lors du tests WB.
Se référer à la notice du kit LDBIO Diagnostics.

RÉACTIFS FOURNIS

1 flacon de lavage concentré (60mL ou 250mL), 1 notice d'utilisation abrégée

MATÉRIEL NÉCESSAIRE MAIS NON FOURNI

Se référer à la notice du kit LDBIO Diagnostics.

CONDITIONS DE CONSERVATION ET PÉREMPTION

Conservation entre 2°C et 8°C. Le réactif est stable jusqu'à la date de péremption indiquée sur l'étiquette du flacon, en l'absence de contamination.
Ne pas utiliser de réactif contaminé ou trouble.

PRÉCAUTIONS D'EMPLOI

Se référer à la notice du kit LDBIO Diagnostics.

Il est préférable d'utiliser des réactifs du même lot. En cas d'utilisation de réactifs de lots différents en assurer la traçabilité.

EUH 210 Fiche de données de sécurité disponible sur demande ainsi que sur notre site internet

www.ldbiodiagnostics.com

PRÉLÈVEMENT DES ÉCHANTILLONS

Se référer à la notice du kit LDBIO Diagnostics.

PRÉPARATION DES RÉACTIFS

Se référer à la notice du kit LDBIO Diagnostics.

MODE OPÉRATOIRE

Se référer à la notice du kit LDBIO Diagnostics.

CONTRÔLE QUALITÉ ET INTERPRÉTATION

Se référer à la notice du kit LDBIO Diagnostics.

LIMITES DU TEST

Se référer à la notice du kit LDBIO Diagnostics.

Le diagnostic d'une maladie infectieuse ne peut pas être établi sur la base du résultat d'un seul test.
Dans le but d'établir un diagnostic clinique, les résultats sérologiques doivent être interprétés en fonction des renseignements disponibles (ex : épidémiologie, clinique, imagerie, biologie...). Ils ne doivent pas faire poser le diagnostic sur leur seule positivité.

PERFORMANCES

Se référer à la notice du kit LDBIO Diagnostics.

REPRODUCTIBILITE

Reproductibilités inter-séries et inter-lots ont été testées. Dans les deux cas, la corrélation sérum à sérum vis-à-vis des bandes spécifiques est excellente.

INTERFERENCES

Se référer à la notice du kit LDBIO Diagnostics.

RÉSOLUTION DE PROBLÈMES

Se référer à la notice du kit LDBIO Diagnostics.

BIBLIOGRAPHIE

Se référer à la notice du kit LDBIO Diagnostics.

R5 - CONCENTRATED WASH BUFFER



#WB-LC60

#WB-LC250

INTENDED USE

Refer to the instructions for use of the LDBIO Diagnostics kit.
This reagent must not be used alone.

PRINCIPLE OF THE TEST / CONDUCT OF THE TEST

Western Blot technique: Wash buffer is used for washing the strips during the WB assay.
Refer to the instructions for use of the LDBIO Diagnostics kit.

REAGENTS SUPPLIED

1 vial of 60 mL or 250mL of wash buffer, 1 abbreviated instruction for use

ADDITIONAL MATERIAL REQUIRED BUT NOT PROVIDED

Refer to the instructions for use of the LDBIO Diagnostics kit.

STORAGE AND STABILITY

Store between 2 and 8°C. The reagent is stable until the expiry date indicated on the outer box and the vial labels.
Do not use contaminated or cloudy reagent.

PRECAUTION FOR USE

Refer to the instructions for use of the LDBIO Diagnostics kit.
It is preferable to use reagents from the same batch. If reagents from different lots are used, ensure traceability.

EUH 210 Safety data sheet available on request and on our website www.ldbiodiagnostics.com

SPECIMEN COLLECTION

Refer to the instructions for use of the LDBIO Diagnostics kit.

PREPARATION OF REAGENTS

Refer to the instructions for use of the LDBIO Diagnostics kit.

TEST PROCEDURE

Refer to the instructions for use of the LDBIO Diagnostics kit.

QUALITY CONTROL AND INTERPRETATION

Refer to the instructions for use of the LDBIO Diagnostics kit.

LIMITATION OF USE

Refer to the instructions for use of the LDBIO Diagnostics kit.

The diagnosis of an infectious disease cannot be established based on a single test result.

Serological results must be interpreted according to available information (e.g., epidemiology, clinical, imaging, biology...) to establish a diagnosis. They should not be used to make a diagnosis based on their positivity alone.

PERFORMANCES

Refer to the instructions for use of the LDBIO Diagnostics kit.

REPRODUCIBILITY

Inter-series and inter-lot reproducibility were tested. In both cases, the serum-to-serum correlation with respect to specific bands is excellent.

INTERFERENCE

Refer to the instructions for use of the LDBIO Diagnostics kit.

TROUBLESHOOTING

Refer to the instructions for use of the LDBIO Diagnostics kit.

BIBLIOGRAPHY

Refer to the instructions for use of the LDBIO Diagnostics kit.

NOTIFICATION DE CHANGEMENT DE VERSION – A lire attentivement

UPDATE NOTIFICATION – Please read carefully

| DATE DE VERSION RELEASE DATE | VERSION | RÉSUMÉ DE LA MODIFICATION MODIFICATION SUMMARY |
|---------------------------------|---------|---|
| 16/08/2022 | Vs 01 | Création / Conception |



NF EN ISO 13485

24 av. Joannes Masset – 69009 LYON – FRANCE
 Tel : +33(0)4 7883 3487 – Fax : +33(0)4 7883 3430
www.ldbiodiagnostics.com – info@ldbiodiag.com